





Test Report SL52025258458701TX Date:June 19,2020 Page 1 of 5

GUANGDONG GOODFEELING HYGIENE MATERIAL TEC CO., LTD NO.10 PHOENIX AVENUE ECO-INDUSTRIAL PARK, DANZAO TOWN, NANHAI, FOSHAN CITY, GUANGDONG PROVINCE, CHINA

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)Medical surgical mask(Non-sterile) (Claimed Type IIR)

Style No. : 18*9.5cm

Composition : (A)Spunbond fabric, Melt-blown fabric

Sample Color : (A)Blue/white

Manufacturer : GUANGDONG GOODFEELING HYGIENE MATERIAL TEC CO., LTD

Country of Destination : United States, EUR

Supplier : GUANGDONG GOODFEELING HYGIENE MATERIAL TEC CO., LTD

Lot No./Batch No. : Not provided

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : May 25, 2020

Testing Period : May 25, 2020 - Jun 19, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the

sample(s) tested, for further details, please refer to the following page(s).



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions/Terms

3rd Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 t (86-21) 61402666







Test Report

SL52025258458701TX

Date:June 19,2020

Page 2 of 5

Comment:

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods	(A)
Clause 5.2 Performance Requirement	
Clause 5.2.2 Bacterial filtration efficiency (BFE)	M
Clause 5.2.3 Breathability	M
Clause 5.2.4 Splash Resistance	M
Clause 5.2.5 Microbial Cleanliness	M
Clause 5.2.6 Biocompatibility	EXCLUDED

Remark: M=Meet EN 14683:2019+AC:2019 Performance Requirement (Type IIR) F=Below EN 14683:2019+AC:2019 Performance Requirement (Type IIR)

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Sara Guo (Account Executive)

Dongjing Liu (Authorized Signatory)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions/Terms







Test Report

SL52025258458701TX

Page 3 of 5

Test Result

EN 14683:2019+AC:2019 Medical Face Masks-Requirements and Test Methods

Clause 5.2 Performance Requirement

Clause 5.2.2 Bacterial filtration efficiency (BFE)*

(EN 14683:2019 Annex B)

Sample: A

1# 2# 3# 4# 5# (BFE), % 99.9 >99.9 99.9 99.9 >99.9

Remark: Performance Requirement: Type I≥95%, Type II≥98%, Type IIR≥98%

Clause 5.2.3 Breathability

(EN 14683 :2019+AC:2019 Annex C)

Sample: A

Test number and location 5 random areas for each specimen (face mask) **Conditioning Parameters** Minimum of 4 hours at 21±5°C and 85±5% R.H.

Test Area 4.9 cm² Flow Rate 8 I/min

1# 2# 3# 4# 5# Differential pressure 34 33 31 34 31 △P (Pa/cm²)

Remark: Performance Requirement: Type I<40 Pa/cm², Type II<40 Pa/cm², Type IIR<60 Pa/cm²



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions/Terms

3rdBuilding, No. 889, Yishan Road, Xuhui District Shanghai, China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 t (86-21) 61402666

f (86-21) 64958763

e sgs.china@sgs.com

^{*} This test standard is not within the accredited scope in SGS Shanghai testing centre, it is carried out by external laboratory accredited by CMA (China Metrology Accreditation).







Date: June 19,2020

中国认可 国际互认 检测 TESTING CNAS L0599

Test Report

SL52025258458701TX

Page 4 of 5

Clause 5.2.4 Splash Resistance

(ISO 22609 :2004, Pressure 16.0 kPa)

Penetration on inside surface								
1#	2#	3#	4#	5#	6#	7#	8#	
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	
9#	10#	11#	12#	13#	14#	15#	16#	
Pass	Pass	Pass	Fail	Pass	Pass	Pass	Pass	
17#	18#	19#	20#	21#	22#	23#	24#	
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass	
25#	26#	27#	28#	29#	30#	31#	32#	
Pass	Pass	Fail	Pass	Pass	Pass	Pass	Pass	
Number of Pass:			30					
Overall result:			Acceptable					

Remark:

- 1) Performance Requirement Type I: N/A, Type II: N/A, Type IIR: ≥16.0kPa
- 2) Distance of the medical face mask target area surface to the tip of cannula is 300±10mm.
- 3) Condition and Test temperature (21±5)° C, relative humidity (85±10)%
- 4) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results

Clause 5.2.5 Microbial Cleanliness

(EN 14683:2019+AC:2019 Annex D and EN ISO 11737-1:2018)

1# 2# 3# 4# 5# CFU/g 5 7 6 7 4

Remark: Performance Requirement: Type I≤30 CFU/g, Type II≤30 CFU/g, Type IIR≤30 CFU/g



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions/Terms

3[™]Building,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国・上海・徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 t (86-21) 61402666

f (86-21) 64958763 f (86-21) 64958763 www.sgsgroup.com.cn e sgs.china@sgs.com








Test Report

SL52025258458701TX

Date:June 19,2020

Page 5 of 5

Sample Photo



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions/Terms

3rdBuilding,No.889,Yishan Road,Xuhui District Shanghai,China 200233 中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t (86-21) 61402666 t (86-21) 61402666 f (86-21) 64958763 f (86-21) 64958763 www.sgsgroup.com.cn e sgs.china@sgs.com